



AURELIA[®] IGNITE

7mil, Nitrile Powder Free Examination Gloves

Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1184: 2002. Materials and articles in contact with foodstuffs.


AURELIA[®] GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.co.uk
+441733 615215

Supermax Healthcare Ltd,
Unit 12-16 Titan Drive
Fengate, Peterborough
PE1 5XN

aureliagloves.co.uk



M	L	*XL	*XXL
97887	97888	97889	97880




Sizes & Dimensions

Single wall thickness

Physical Properties

M 95+10 mm	min. 240 mm
L 110+10 mm	min. 240 mm
XL 120+10 mm	min. 240 mm
XXL > 120 mm	min. 240 mm



- ① min. 0.25 mm
- ② min. 0.20 mm
- ③ min. 0.09 mm

	Unaged	Aged
Tensile Strength	Min. 14.0 MPa	Min. 14.0 MPa
Elongation	Min. 500%	Min. 400%
Strength at Break	Min. 6N	Min. 6N



Nitrile



Powder Free



5 Years



Diamond



8.4g +/- 0.2g



Orange

