



AURELIA® BOLD MAX

6mil, Nitrile Powder Free Examination Gloves

Regulations	PPE Regulation (EU) 2016/425 Medical Device Regulation (EU) 2017/745, 93/42/EEC
Quality Standards	ISO 9001:2015 Quality Management Systems ISO 13485: 2016 Medical Devices Quality Management Systems
Harmonised Standards	BS EN ISO 374 - 1,2,4,5 Protective gloves against dangerous chemicals and microorganisms. BS EN 455 1-4 Medical gloves for single use. EN 420: 2003 + A1: 2009 Protective gloves. General requirements and test methods. BS EN 1184: 2002. Materials and articles in contact with foodstuffs.

AURELIA® GLOVES

A SUPERMAX HEALTHCARE BRAND

sales@supermax.co.uk

+441733 615215

Supermax Healthcare Ltd,
Unit 12-16 Titan Drive
Fengate, Peterborough
PE1 5XN

aureliagloves.co.uk



10 x 50

M L XL XXL

9789A7 9789A8 9789A9 9789A0



Sizes & Dimensions	Single wall thickness	Physical Properties
		
M 95+10 mm	min. 240 mm	① min. 0.20 mm
L 110+10 mm	min. 240 mm	② min. 0.16 mm
XL 120+10 mm	min. 240 mm	③ min. 0.08 mm
XXL > 120 mm	min. 240 mm	
		Unaged Aged
		Tensile Strength Min. 14.0 MPa Min. 14.0 MPa
		Elongation Min. 500% Min. 400%
		Strength at Break Min. 6N Min. 6N



Nitrile



Powder Free



5 Years



Diamond



7.5g +/- 0.2g



Black

